

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

SEP 23 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130297</u> ✓
Name <u>JAMES C. HOELSCHER</u>		ID# <u>5343</u> Date <u>9-21-09</u> ✓
A	Agency <u>Hooper Bay Police Dept.</u>	Phone # <u>758-4615</u>
Instrument Location <u>Hooper Bay Police Dept. Hooper Bay, AK</u> ✓		
Alco S/N <u>X301526</u> ✓ Target Value <u>1.080</u> ✓ High Pressure <u>800 PSI</u>		
B	Alco Test Value Average <u>1.081</u> ✓ 1 <sup>st</sup> Alco 2 <sup>nd</sup> Alco <u>1.080</u> ✓	
Signature <u>[Signature]</u> ✓		
(OVER)		

BMB  
10/13/09

(Do Not write in the area below)


- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
  - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
  - (3) I am the Scientific Director of the State Breath Alcohol Program.
  - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
  - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
  - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 20<sup>th</sup> day of October, 2009.

[Signature] (Notary Seal Stamp)

Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm breath Test Instrument*

*State of Alaska*

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

DataMaster cdm S/N

130297 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

8MB  
10/13/09

8MB  
10/13/09